



SMITH RIVER BIOLOGICALS

Quality Biologicals And Prepared Media For The Laboratory Professional

QUALITY ASSURANCE CERTIFICATE

11-SDACO SABOURAUD DEXTROSE AGAR with CHLORAMPHENICOL 20mL

Lot: 1812280

Exp. 3/28/19

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:

PERFORMANCE TESTING

<u>Test Organisms</u>	<u>Growth</u>
<i>Escherichia coli</i> ATCC® 25922	Inhibited
<i>Trichophyton interdigitale</i> ATCC® 9533	4+
<i>Candida albicans</i> ATCC® 10231	4+
<i>Aspergillus brasiliensis</i> ATCC® 16404	4+

Growth approved by: _____

(QA Manager)

MICROBIAL LOAD

Reports are based on U.S.P.XXI(1211)

Lot size: 5L Number of units sampled: 10

Number of units failed at 35°C and ambient temperature: 0

Samples tested at 35°C and ambient temperature.

Final pH at 25°C: 5.6±0.3

Lot approved by: _____

(QA Manager)

Not intended for IVD use

THIS DOCUMENT MEETS THE STANDARDS FOR INSPECTION AND ACCREDITATION CURRENTLY ESTABLISHED BY THE COLLEGE OF AMERICAN PATHOLOGIST, NCCLS-M22T, U.S.P.XXI(1211) AND MS105D.