



SMITH RIVER BIOLOGICALS
Quality Biologicals And Prepared Media For The Laboratory Professional

Certificate of Analysis

Lethen Broth

Product Number: 22-LeB9.9
Lot Number: 1908190

Fill Volume: 9.9mL
Expiration: 2/19/2020

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
<i>E. coli</i> ATCC® 25922	Excellent
<i>S. aureus</i> ATCC® 25923	Excellent
<i>M. luteus</i> ATCC® 9341	Excellent
<i>A. brasiliensis</i> ATCC® 16404	Excellent
<i>C. albicans</i> ATCC® 10231	Excellent
<i>B. subtilis</i> ATCC® 6633	Excellent

Microbial testing was performed aerobically at 35°C for 18-24 hours

Growth approved by: _____
(QA Manager)

Microbial Load

Reports are based on USPXXI(1211)

Lot Size:	3L
Number of Units Sampled:	12
Final pH at 25°C:	7.0±0.2

Appearance

Uninoculated Lethen Broth is dark yellow/tan and somewhat hazy

Lot approved by: _____
(QA Manager)

THIS DOCUMENT MEETS THE STANDARDS FOR INSPECTION AND ACCREDITATION CURRENTLY ESTABLISHED BY THE COLLEGE OF AMERICAN PATHOLOGIST, NCCLS-M22T, U.S.P.XXI(1211) AND MS105D. NOT FOR IVD USE.

