SMITH RIVER BIOLOGICALS

Quality Biologicals And Prepared Media For The Laboratory Professional

Certificate of Analysis

TSA with Lecithin and Polysorbate 80

Product Number: 24-MCT20 Fill Volume: 15mL

Lot Number: 1908050 Expiration: 2/5/2020

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
E. coli ATCC©25922	Good Growth
S. aureus ATCC©25923	Good Growth
M. luteus ATCC©9341	Good Growth
A. brasiliensis ATCC©16404	Good Growth
C. albicans ATCC©10231	Good Growth
B. subtilis ATCC©6633	Good Growth

Microbial testing was performed aerobically at 35°C for 18-24 hours

Growth approved

(QA Manager)

Microbial Load

Lot Size:	3L
Number of Units Sampled:	16
Final pH at 25°C:	7.3±0.2

Appearance

Prepared TSA with Lecithin and Polysorbate 80 plates are light to dark tan and opaque

Lot approved

(QA Manager)

Not for IVD use.