SMITH RIVER BIOLOGICALS

Quality Biologicals And Prepared Media For The Laboratory Professional

Certificate of Analysis

Mueller Hinton Agar

Product Number: 11-MH5%A
Lot Number: 1910220

Fill Volume: 18 ml
Expiration: 2/22/2020

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
**S. pneumoniae* ATCC©49619	Growth
E. coli ATCC©25922	Growth
S. aureus ATCC©25923	Growth
E. faecalis ATCC©29212	Growth

**Organism recommended by CLSI M02-A*

Growth approved by:_____ _____
(QA Manager)

Microbial Load

Lot Size:	1.5L
Number of Units Sampled:	28
Final pH at 25°C:	7.3±0.2

Appearance

Prepared Mueller Hinton Blood Agar is bright to dark red and turbid

Lot approved by:_____
(QA Manager)

Not for IVD use.