



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

PPLO Agar

Product Number: 18-PPLOA

Lot Number: 2007020

Expiration: 10/2/2020

Fill Volume: 10mL

Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
<i>E. coli</i> ATCC® 25922	Inhibited
<i>C. albicans</i> ATCC® 10231	Inhibited
<i>S. aureus</i> ATCC® 25923	Inhibited
<i>M. pneumoniae</i> ATCC® 15531	Growth

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size:	1.5L
Number of Units Sampled:	27
Final pH at 25°C:	7.8±0.2

Appearance

Uninoculated PPLO agar is amber and slightly hazy

Lot approved by: _____

(QA Manager)

Custom Formulation

