



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis Lethen Broth

Product Number: XX-LeB90
Lot Number: 2007090


Fill Volume:
Expiration: 1/9/21

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and the following results and reactions:

Performance Testing

| <u>Test Organisms</u> | <u>Growth</u> |
|--------------------------------|---------------|
| <i>E. coli</i> ATCC® 25922 | Excellent |
| <i>S. aureus</i> ATCC® 25923 | Excellent |
| <i>C. albicans</i> ATCC® 10231 | Excellent |

Microbial testing was performed anaerobically at 35-37°C for 18-24 hours
Sterility check was performed at 35-37°C for 24 hours

Growth approved by: 
(QA Manager)

Microbial Load

Lot Size: 30
Number of Units Sampled: 20
Final pH at 25°C: 7.0±0.2

Appearance

Uninoculated Lethen Broth is dark yellow/tan and somewhat hazy

Lot approved by: 
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.

S
onal

90 mL

re
gave the