



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis Anaerobic Blood Agar

Product Number: 11-ABA00
Lot Number: 2008060

Expiration: 11/6/2020


Fill Volume: 18mL
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
<i>F. nucleophorum</i> ATCC® 25286	Growth
<i>P. anaerobius</i> ATCC® 27337	Growth
<i>S. aureus</i> ATCC® 25923	Growth at 72h

Microbial testing was performed anaerobically at 35-37°C for 18-24 hours
Sterility check was performed at 35-37°C for 24 hours

Growth approved by: 
(QA Manager)

Microbial Load

Lot Size: 10L
Number of Units Sampled: 29
Final pH at 25°C: 7.3±02

Appearance

Uninoculated media is dark red to red brown and opaque

Lot approved by: 
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.