



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Cystine Tellurite Blood Agar

Product Number: 11-CTBA0
Lot Number: 2008100

Expiration: 11/10/2020

Fill Volume: 20 mL
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Reaction</u>
<i>C. diphtheriae</i> ATCC® 13812	Growth	Black
<i>S. sanguis</i> ATCC® 10556	Inhibited	NA
<i>S. aureus</i> ATCC® 25923	Inhibited	NA

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility check was performed at 35-37°C for 24 hours

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size: 1L
Number of Units Sampled: 29
Final pH at 25°C: 7.4±0.2

Appearance

Uninoculated Cystine Tellurite Blood Agar is bright to dark red and opaque

Lot approved by: _____

(QA Manager)

This product does is not intended for the diagnosis or treatment of human disease.