



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

A-1 Medium, Single Strength

Product Number: 22-A1S10
Lot Number: 2009100

Fill Volume: 10 ml
Expiration: 3/10/2020

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Reaction</u>
<i>E. coli</i> ATCC®25922	Growth	Gas present
<i>P. mirabilis</i> ATCC®12453	Growth	Gas absent
<i>P. aeruginosa</i> ATCC®27853	Inhibited	NA

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: 
(QA Manager)

Microbial Load

Lot Size: 6L
Number of Units Sampled: 12
Final pH at 25°C: 6.9±0.2

Appearance

Prepared A-1 Medium is light amber and opalescent

Lot approved by: 
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.