



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis Cetrimide/MSA/TSA Triplates

Product Number: 13-CetMSATSA
Lot Number: 2012020

Expiration: 3/2/2021

Fill Volume: 6mL/side
Fill Depth:

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Cetrimide</u>	<u>MSA</u>	<u>TSA</u>
<i>E. coli</i> ATCC® 25922	Inhibited	Inhibited	Excellent
<i>P. aeruginosa</i> ATCC® 27853	Good growth Green colonies	Inhibited	Excellent
<i>P. mirabilis</i> ATCC® 12453	Inhibited	Excellent <i>Mannitol -</i>	Excellent
<i>S. aureus</i> ATCC® 25923	Inhibited	Excellent <i>Mannitol +</i>	Excellent
<i>E. faecalis</i> ATCC® 29212	Inhibited	Inhibited	Excellent

Microbial testing was performed aerobically at 35°C for 18-24 hours

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size: 30 Liters
Number of Units Sampled: 27
Final pH at 25°C: 7.2±0.2; MSA: 7.4±0.2; TSA: 7.3±0.2

Lot approved by: _____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.