



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Rappaport Vassiliadis Enrichment Broth

Product Number: 22-RVS10

Fill Volume: 10 ml

Lot Number: 2102020

Expiration: 8/2/2021

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

| <u>Test Organisms</u> | <u>Growth</u> |
|-------------------------------------|---------------|
| ** <i>S. typhimurium</i> ATCC®14028 | Excellent |
| ** <i>S. aureus</i> ATCC®6538 | Excellent |

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

**Organisms required by USP <62> Microbiological Tests

Growth approved by: _____

(QA Manager)

Microbial Load

| | |
|--------------------------|---------|
| Lot Size: | 0.7L |
| Number of Units Sampled: | 18 |
| Final pH at 25°C: | 5.2±0.2 |

Appearance

RVS Broth is dark blue to teal and opalescent. Some precipitate may be present

Lot approved by: _____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.