SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Letheen Agar, Modified

Product Number: B-LeB250 Fill Volume: 250 mL
Lot Number: 2104260 Expiration: 10/26/2021

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
E. coli ATCC® 25922	Excellent
S. aureus ATCC® 25923	Excellent
C. albicans ATCC® 10231	Excellent

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____ _____
(QA Manager)

Microbial Load

Lot Size:	11.5L
Number of Units Sampled:	1
Final pH at 25°C:	7.2 ± 0.2

Appearance

Modified Letheen Agar is yellowish tan and opaque.

Lot approved by:_____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.