



# SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

## Certificate of Analysis Potato Dextrose Agar

Product Number: 11-PDA00  
Lot Number: 2105100

Expiration: 8/10/2021

Fill Volume: 20 ml  
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

### **Performance Testing**

<u>Test Organisms</u>	<u>Growth</u>
<i>A. brasiliensis</i> ATCC® 16404	Growth
<i>C. albicans</i> ATCC® 10231	Growth
<i>T. interdigitale</i> ATCC® 9533	Growth
<i>E. coli</i> ATCC® 25922	Growth

Microbial testing was performed aerobically at 35-37°C for 18-24 hours  
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: \_\_\_\_\_  
(QA Manager)

### **Microbial Load**

Lot Size: 5L  
Number of Units Sampled: 30  
Final pH at 25°C: 5.6 ±0.2

### **Appearance**

PDA is light amber and opalescent

Lot approved by: \_\_\_\_\_  
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.