



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Tryptic Soy Agar with 5% Sheep Blood Slants

Product Number: 22-TSAB08

Fill Volume: 8 ml

Lot Number: 2105120

Expiration 11/12/2021

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:


Performance Testing

| <u>Test Organisms</u> | <u>Growth</u> |
|-----------------------------------|-------------------------|
| * <i>S. aureus</i> ATCC®25923 | Growth |
| * <i>S. agalactiae</i> ATCC®12386 | Growth |
| * <i>S. pyogenes</i> ATCC®19615 | Growth, beta hemolysis |
| * <i>S. pneumoniae</i> ATCC®6305 | Growth, alpha hemolysis |
| * <i>E. coli</i> ATCC®25922 | Growth |

*Organisms required by CLSI M22-A3

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: 


(QA Manager)

Microbial Load

| | |
|--------------------------|---------|
| Lot Size: | 1L |
| Number of Units Sampled: | 15 |
| Final pH at 25°C: | 7.3±0.2 |

Appearance

Uninoculated slants are bright red to dark red and opaque

Lot approved by: 

(QA Manager)

This product is not intended for diagnosis or treatment of human disease.