



# SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

## Certificate of Analysis

### Burkholderia Cepacia Selective Agar

Product Number: 13-BCSA0  
Lot Number: 2110190

Expiration: 1/19/2022

Fill Volume: 6 mL/side  
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

#### Performance Testing

##### Test Organisms

*Pseudomonas aeruginosa* ATCC®27853

*Staphylococcus aureus* ATCC®25923

*Enterococcus faecalis* ATCC®29212

##### Growth

Partial to Complete Inhibition

Partial to Complete Inhibition

Partial to Complete Inhibition

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: Olga Ald  
(QA Manager)

#### Microbial Load

Lot Size: 3L

Number of Units Sampled: 30

Final pH at 25°C: 7.0 ± 0.2

#### Appearance

Prepared agar is pink-purple and clear with no precipitate

Lot approved by: Olga Ald  
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.