



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Burkholderia Cepacia Selective Agar

Product Number: 13-BCSA0
Lot Number: 2201100

Expiration: 4/10/2022

Fill Volume: 6 mL/side
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms

Pseudomonas aeruginosa ATCC®27853

Staphylococcus aureus ATCC®25923

Enterococcus faecalis ATCC®29212

Growth

Partial to Complete Inhibition

Partial to Complete Inhibition

Partial to Complete Inhibition

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____


(QA Manager)

Microbial Load

Lot Size: 5L

Number of Units Sampled: 30

Final pH at 25°C: 7.0 ± 0.2

Appearance

Prepared agar is pink-purple and clear with no precipitate

Lot approved by: _____


(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.