



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis Standard Methods Agar

Product Number: 11-SMA00

Lot Number: 2206080

Expiration: 9-8-22

Fill Volume: 20 mL

Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
<i>E. coli</i> ATCC®25922	Excellent
<i>S. epidermidis</i> ATCC®12228	Excellent
<i>C. albicans</i> ATCC®10231	Excellent
<i>B. subtilis</i> MCM 3010	Excellent
<i>P. aeruginosa</i> ATCC®27853	Excellent

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size: 70L

Number of Units Sampled: 30

Final pH at 25°C: 7.0±0.2

Appearance

Uninoculated agar is light to medium amber and clear to hazy.

Lot approved by: _____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.