



# SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

## Certificate of Analysis

### Columbia Agar with 5% Sheep Blood

Product Number: 11-CA5%0

Expiration: 12/31/2022

Fill Volume: 18mL

Lot Number: 2208310

Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

#### **Performance Testing**

<u>Test Organisms</u>	<u>Growth</u>
<i>L. monocytogenes</i> ATCC® 7644	Growth
<i>S. pyogenes</i> ATCC® 19615	Growth
<i>E. coli</i> ATCC® 25922	Growth
<i>S. aureus</i> ATCC® 25923	Growth

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: \_\_\_\_\_

(QA Manager)

#### **Microbial Load**

Lot Size:	10L
Number of Units Sampled:	30
Final pH at 25°C:	7.3±0.2

#### **Appearance**

Uninoculated media is medium to dark red and opaque

Lot approved by: \_\_\_\_\_

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.