



# **SMITH RIVER BIOLOGICALS**

*Quality Biologicals and Prepared Media for the Laboratory Professional*

## **Certificate of Analysis**

### **Pseudomonas Isolation Agar**

Product Number: 11-PIA00  
Lot Number: 2211010

Expiration: 2/1/2023

Fill Volume: 20 ± 2 mL  
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

#### **Performance Testing**

<u>Test Organisms</u>	<u>Growth</u>
<i>E. coli</i> ATCC®25922	Inhibited
<i>P. mirabilis</i> ATCC®12453	Inhibited
<i>P. aeruginosa</i> ATCC®27583	Growth, green colonies

Microbial testing was performed aerobically at 35-37°C for 18-24 hours  
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:   
(QA Manager)

#### **Microbial Load**

Lot Size: 3L  
Number of Units Sampled: 30  
Final pH at 25°C: 7.0±0.2

#### **Appearance**

Pseudomonas Isolation Agar is light amber and slightly turbid

Lot approved by:   
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.