



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Sabouraud Dextrose Agar Triplates

Product Number: 13-SDA00

Fill Volume: 6 mL / side

Lot Number: 2210310

Fill Depth: 3-5 mm

Expiration: 1-31-23

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms

A. brasiliensis ATCC®16404

C. albicans ATCC®10231

T. interdigitale ATCC®9533

E. coli ATCC®25922

Growth

Growth after 1-5 days

Growth after 1-2 days

Growth after 3-7 days

Growth after 24 hours

Microbial testing was performed aerobically at 20-25°C

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size:

4L

Number of Units Sampled:

30

Final pH at 25°C:

5.6±0.2

Appearance

Prepared SDA is light to dark amber

Lot approved by: _____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.