



# **SMITH RIVER BIOLOGICALS**

Quality Biologicals and Prepared Media for the Laboratory Professional

## **Certificate of Analysis** **Sabouraud Dextrose Agar**

Product Number: 11-SDA00  
Lot Number: 2212120

Expiration: 3/12/23

Fill Volume: 20 ± 2 mL  
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

### **Performance Testing**

<u>Test Organisms</u>	<u>Growth</u>
<i>A. brasiliensis</i> ATCC®16404	Growth after 1-5 days
<i>C. albicans</i> ATCC® 10231	Growth after 1-2 days
<i>T. interdigitale</i> ATCC®9533	Growth after 3-7 days
<i>E. coli</i> ATCC®25922	Growth after 24 hours

Microbial testing was performed aerobically at 20-25°C  
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: \_\_\_\_\_

(QA Manager)

### **Microbial Load**

Lot Size: 2.5L  
Number of Units Sampled: 30  
Final pH at 25°C: 5.6±0.2

### **Appearance**

Prepared SDA is light to dark amber

Lot approved by: \_\_\_\_\_

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.