



# SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

## Certificate of Analysis

### Baird Parker Agar

Product Number: 11-BPA00  
Lot Number: 2305090

Expiration: 8/9/2023

Fill Volume: 20 ± 2 mL  
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

#### Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Reaction</u>	<u>Lecithianase</u>
<i>E. coli</i> ATCC® 25922	Inhibited	NA	NA
<i>P. mirabilis</i> ATCC® 12453	4+	Brown	(-)
<i>S. aureus</i> ATCC® 25923	4+	Black	(+)
<i>S. epidermidis</i> ATCC® 12228	4+	Black	(+)
<i>S. aureus</i> ATCC® 6538	4+	Black	(+)

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: Olga Ald  
(QA Manager)

#### Microbial Load

Lot Size: 3L  
Number of Units Sampled: 30  
Final pH at 25°C: 7.0±0.2

#### Appearance

Uninoculated media is pale yellow and opaque, may contain yellow flakes

Lot approved by: Olga Ald  
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.