



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

EC Medium

Product Number: 22-ECS10

Lot Number: 2305170

Fill Volume: 10 ± 0.2 mL

Expiration: 11/17/2023

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Gas Production</u>
<i>E. coli</i> ATCC® 25922	Good	+
<i>E. faecalis</i> ATCC® 29212	Inhibited	NA

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____
(QA Manager)

Microbial Load

Lot Size: 0.8L
Number of Units Sampled: 6
Final pH at 25°C: 6.9±0.2

Appearance

Uninoculated media is amber and clear

Lot approved by: _____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.