



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis mEndo Agar

Product Number: 11-mEndo
Lot Number: 2305250

Fill Volume: 20 ± 2 mL
Expiration: 11/25/2023

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Results</u>
<i>E. coli</i> ATCC®25922	Good	Green with metallic sheen
<i>E. aerogenes</i> ATCC®13048	Good	Green with metallic sheen
<i>S. typhimurium</i> ATCC®14028	Good	Pink or red
<i>S. aureus</i> ATCC®25923	Inhibited	NA

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____
(QA Manager)

Microbial Load

Lot Size: 6.2L
Number of Units Sampled: 30
Final pH at 25°C: 7.2±0.2

Appearance

Lot approved by: _____
(QA Manager)

THIS DOCUMENT MEETS THE STANDARDS FOR INSPECTION AND ACCREDITATION CURRENTLY ESTABLISHED BY THE COLLEGE OF AMERICAN PATHOLOGIST, NCCLS-M22T, U.S.P.XXI(1211) AND MS105D. NOT INTENDED FOR THE DIAGNOSIS OR TREATMENT OF HUMAN DISEASE.