



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

TSA with Lecithin and Polysorbate 80

Product Number: 17-MCT00

Fill Volume: 15 ml

Lot Number: 2306070

Expiration: 10/7/2023

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
<i>E. coli</i> ATCC@25922	Good Growth
<i>S. aureus</i> ATCC@25923	Good Growth
<i>C. albicans</i> ATCC@10231	Good Growth

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: 

(QA Manager)

Microbial Load

Lot Size:	7L
Number of Units Sampled:	30
Final pH at 25°C:	7.3±0.2

Appearance

Prepared TSA with Lecithin and Polysorbate 80 plates are light to dark tan and opaque

Lot approved by: 

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.