

Certificate of Analysis

Burkholderia Cepacia Selective Agar

Product Number: 11-BCSA0 Expiration: 10/6/2023 Fill Volume: $20 \text{ mL} \pm 2$ Lot Number: 2307060 Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Growth

Partial to Complete Inhibition

Partial to Complete Inhibition

Partial to Complete Inhibition

Performance Testing

P. aeruginosa ATCC®27853

S. aureus ATCC®25923

E. faecalis ATCC®29212

Test Organisms

Sterility testing was performed at 35-37°C for 24 hours	
Growth approved by:	Ole ald
	(QA Manager)
Microbial Load	
Lot Size:	5L
Number of Units Sampled:	30
Final pH at 25°C:	7.0 ± 0.2
<u>Appearance</u>	
Prepared agar is pink-purple and clear with no precipitate	
Lot approved by:	Ole ald
	(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.

Microbial testing was performed aerobically at 35-37°C for 18-24 hours