



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Brilliant Green Agar with Novobiocin

Product Number: 11-BGAN

Expiration: 10/11/2023

Fill Volume: 20 mL±2 ml

Lot Number: 2307110

Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms

Growth

S. typhimurium ATTC® 14028

Excellent

E. coli ATTC® 25922

Inhibited

S. enterica ATTC® 14028

Fair/Good

S. aureus ATTC® 25923

Inhibited

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size:

5 L

Number of Units Sampled:

30

Final pH at 25°C:

6.9 ± 0.2

Appearance

Prepared plates are a transparent dark green.

Lot approved by: _____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.