



# SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

## Certificate of Analysis

### Brilliance Bacillus cereus Agar

Product Number: 16-BBCA0

Expiration: 9/26/2023

Fill Volume: 60 mL

Lot Number: 2306260

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

#### Performance Testing

##### Test Organisms

*B.subtilis* ATCC® 6633

*P. aeruginosa* ATCC® 27853

*E. coli* ATCC® 25922

*E. faecalis* ATCC® 29212

*B.cereus* ATCC® 13061

##### Growth

Inhibited

Partial to Complete Inhibition

Partial to Complete Inhibition

Inhibited

Growth

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: \_\_\_\_\_

  
(QA Manager)

#### Microbial Load


Lot Size: 2L

Number of Units Sampled: 12

Final pH at 25°C: 7.2 ± 0.2

#### Appearance

Lot approved by: \_\_\_\_\_

  
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.