



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

TSA with Poly 80 and Lecithin (MCT)

Product Number: 11-MCTA0

Expiration: 10/27/23

Fill Volume: 20 ± 2 mL

Lot Number: 2307270

Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
<i>E. coli</i> ATCC® 25922	Excellent
<i>S. aureus</i> ATCC® 6538	Excellent
<i>C. albicans</i> ATCC® 10231	Excellent

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: Olivia Ald
(QA Manager)

Microbial Load

Lot Size:	5 L
Number of Units Sampled:	30
Final pH at 25°C:	7.3±0.2

Appearance

Uninoculated media is light beige and slightly hazy

Lot approved by: Olivia Ald
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.