



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

MRS (De Man, Rogosa, and Sharpe) Agar

Product Number: 11-MRS000


Expiration: 11/29/2023

Fill Volume: 20 ± 2 mL

Lot Number: 2308290

This is to certify that representative samples of this lot of media were tested for sterility in accordance with routine microbiological procedures in our Quality Assurance Laboratory. Acceptable sterility results were obtained.

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: 
(QA Manager)

Microbial Load

Lot Size: 3L
Number of Units Sampled: 30
Final pH at 25°C: 5.7 ± 0.2

Appearance

Lot approved by: 
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.