



SMITH RIVER BIOLOGICAL

Quality Biologicals and Prepared Media for the Laboratory Profession

Certificate of Analysis

Chocolate Agar with VX Supplement

Product Number: 11-ChocVX

Expiration: 1/13/2024

Fill Volume:

Lot Number: 2309130

Fill Depth:

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
<i>N. gonorrhoeae</i> ATCC® 43069	Growth
<i>H. influenzae</i> ATCC® 10211	Growth

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility check was performed at 35-37°C for 24 hours

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size:	7 L
Number of Units Sampled:	30
Final pH at 25°C:	7.1±0.2

Appearance

Uninoculated media is brown and opaque, mimics milk chocolate

Lot approved by: _____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.

S
onal

18mL
3-5 mm

ine
I gave the