



# **SMITH RIVER BIOLOGICALS**

Quality Biologicals and Prepared Media for the Laboratory Professional

## **Certificate of Analysis** **Mueller Hinton Agar**

Product Number: 16-MHA00

Expiration: 12/29/2023

Fill Volume: 60 mL

Lot Number: 2308290

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. This media passed sterility testing standards.

### **Performance Testing**

<u>Test Organisms</u>	<u>Growth</u>
<i>S. pneumoniae</i> ATCC® 49619	Growth
<i>E. coli</i> ATCC® 25922	Growth

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:   
(QA Manager)

### **Microbial Load**

Lot Size:	64L
Number of Units Sampled:	26
Final pH at 25°C:	7.3±0.2

### **Appearance**

Lot approved by:   
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.