



# SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

## Certificate of Analysis

### EE Mossel Broth

Product Number: XX-EEMB100

Lot Number: 2312040

Fill Volume: 100±2 mL

Expiration: 6/4/2024

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

#### **Performance Testing**

<u>Test Organisms</u>	<u>Growth</u>
<i>S. typhimurium</i> ATCC®14028	Growth
<i>E. coli</i> ATCC®25922	Growth
<i>P. aeruginosa</i> ATCC®9027	Growth
<i>S. aureus</i> ATCC®25923	Inhibited
<i>S. aureus</i> ATCC®6538	Inhibited

Growth in cups indicated by turbidity

Microbial testing was performed aerobically at 35°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: \_\_\_\_\_

  
(QA Manager)

#### **Microbial Load**

Lot Size:	12 L
Number of Units Sampled:	20
Final pH at 25°C:	7.2±0.2

#### **Appearance**

EE Mossel Broth is dark green and clear with no precipitate

Lot approved by: \_\_\_\_\_

  
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.