

Certificate of Analysis

Listeria Medium

Product Number: XX-Lis60 Fill Volume: 60±2 mL
Lot Number: 2401170 Expiration: 7/14/2024

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing Test Organisms Growth L. monocytogenes ATCC® 7644 4+ L. monocytogenes ATCC® 19114 4+ L. monocytogenes ATCC® 19115 Microbial testing was performed aerobically at 35°C for 18-24 hours _____ Ole all____ Growth approved by: (QA Manager) Microbial Load Report is based on U.S.P.XXI(1211) Lot Size: 60 L Number of Units Sampled: 11 **Appearance** Uninoculated media is medium to dark amber and clear, may contain a slight precipitate Lot approved by:____ (QA Manager)

THIS DOCUMENT MEETS THE STANDARDS FOR INSPECTION AND ACCREDITATION CURRENTLY ESTABLISHED BY THE COLLEGE OF AMERICAN PATHOLOGIST, NCCLS-M22T, U.S.P.XXI(1211) AND MS105D. NOT INTENDED FOR THE DIAGNOSIS OR TREATMENT OF HUMAN DISEASE.