

Certificate of Analysis

Listeria Medium

Product Number: XX-Lis60 Lot Number: 2403050 Fill Volume: 60±2 mL Expiration: 9/5/2024

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
L. monocytogenes ATCC® 7644	4+
L. monocytogenes ATCC® 19114	4+
L. monocytogenes ATCC® 19115	4+

Microbial testing was performed aerobically at 35°C for 18-24 hours

Growth approved by:

_____ (QA Manager)

Microbial Load

Report is based on U.S.P.XXI(1211) Lot Size: 60 L Number of Units Sampled: 11

Appearance

Uninoculated media is medium to dark amber and clear, may contain a slight precipitate

Lot approved by:	Ole ald_

(QA Manager)

THIS DOCUMENT MEETS THE STANDARDS FOR INSPECTION AND ACCREDITATION CURRENTLY ESTABLISHED BY THE COLLEGE OF AMERICAN PATHOLOGIST, NCCLS-M22T, U.S.P.XXI(1211) AND MS105D. NOT INTENDED FOR THE DIAGNOSIS OR TREATMENT OF HUMAN DISEASE.