



# SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

## Certificate of Analysis

### Cetrimide Agar

Product Number: 11-Cet00

Expiration: 6/4/2026

Fill Volume: 20 ± 2 mL

Lot Number: 2603040

Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

#### **Performance Testing**

<u>Test Organisms</u>	<u>Growth</u>	<u>Reaction</u>
<i>E. coli</i> ATCC®25922	Inhibited	NA
<i>S. aureus</i> ATCC®25923	Inhibited	NA
<i>P. aeruginosa</i> ATCC®27853	Growth	Green

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: \_\_\_\_\_

(QA Manager)

#### **Microbial Load**

Lot Size: 5L  
Number of Units Sampled: 30  
Final pH at 25°C: 7.2±0.2

#### **Appearance**

Prepared Cetrimide plates are white and cloudy with precipitate

Lot approved by: \_\_\_\_\_

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.