



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Washed Sheep Blood Agar with Mitomycin C and Calcium Chloride

Product Number: 11-WSBAM

Fill Volume: 20 ± 2 mL

Lot Number: 2603100

Fill Depth: 3-5 mm

Expiration: 6/10/2026

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Reaction</u>
<i>E. coli</i> ATCC®25922	Excellent	Hemolysis at 6 hours
<i>E. coli</i> ATCC®35150	Excellent	Hemolysis at 24 hours

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size: 3L
Number of Units Sampled: 26
Final pH at 25°C: 7.2±0.2

Appearance

Uninoculated Washed Sheep Blood Agar is bright to dark red and opaque.

Lot approved by: _____

(QA Manager)

This product is not intended for *in vitro* diagnostic use