



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis **Urea Agar Slants**

Product Number: 22-UAS08
Lot Number: 2602100

Fill Volume: 8 ± 0.2 mL
Expiration: 2/10/2027

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Reaction</u>
<i>E. coli</i> ATCC® 25922	Growth	Negative, Yellow
<i>P. hauseri</i> ATCC® 13315	Growth	Positive, Pink
<i>K. pneumoniae</i> ATCC® 13883	Growth	Positive, Pink
<i>P. mirabilis</i> ATCC® 12435	Growth	Positive, Pink

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____  _____
(QA Manager)

Microbial Load

Lot Size: 14L
Number of Units Sampled: 12
Final pH at 25°C: 6.8±0.2

Appearance

Uninoculated media is pale orange and slightly hazy

Lot approved by: _____  _____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.