SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Violet Red Bile Glucose Agar

Product Number: 11-VRBGA Expiration: 6/26/2026 Fill Volume: 20 ± 2 mL
Lot Number: 2603260 Fill Depth: 3-5mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth	Reaction
E. coli ATCC®25922	Good	Pink to Red with Precipitate
E. faecalis ATCC®29212	Inhibited	NA
k. aerogenes ATCC®13048	Fair to Good	Pink Colonies
S. enterica ATCC®14028	Fair to Good	Colorless Colonies
S. aureus ATCC® 25923	Inhibited	N/A

Microbial testing was performed aerobically at 35°C for 18-24 hours

Growth approved by:_____ _____
(QA Manager)

Microbial Load

Lot Size:	1.5L
Number of Units Sampled:	30
Final pH at 25°C:	7.4±0.2

Appearance

Uninoculated media is light to medium purple and clear, may contain dakr spots

Lot approved by:_____
(QA Manager)

This product is not intended for diagnosis or treatment of human disease.