



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

XLD Agar

Product Number: 11-XLD00

Lot Number: 2604010

Expiration: 7-1-26

Fill Volume: 20 ± 2 mL

Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms

Growth

Reaction

E. coli ATCC®25922

Partial to Complete

Yellow to Red

S. flexneri ATCC®12022

Inhibition

Red

S. typhimurium ATCC®14028

Growth

Red with Black Centers

E. faecalis ATCC®29212

Partial to Complete

Inhibition

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: _____

(QA Manager)

Microbial Load

Lot Size:

1.5L

Number of Units Sampled:

20

Final pH at 25°C:

7.4±0.2

Appearance

Prepared plates are red to orange and slightly cloudy

Lot approved by: _____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.