



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis **Mueller Hinton Agar**

Product Number: 16-MHA00

Expiration: 7/23/2026

Fill Volume: 60 mL

Lot Number: 2603230

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
* <i>S. pneumoniae</i> ATCC®49619	Growth
<i>E. coli</i> ATCC®25922	Growth

*Organism recommended by CLSI M02-A

Microbial testing was performed aerobically at 35-37°C for 18-24 hours

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: 
(QA Manager)

Microbial Load

Lot Size:	103L
Number of Units Sampled:	20
Final pH at 25°C:	7.3±0.2

Appearance tannish

Lot approved by: 
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.