



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis **Sabouraud Dextrose Broth**

Product Number: XX-SDB100

Lot Number: 2604140

Fill Volume: 100 ± 2 mL

Expiration: 10/14/2026


This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
<i>A. brasiliensis</i> ATCC®16404	Growth after 1-5 days
<i>C. albicans</i> ATCC®10231	Growth after 1-2 days
<i>T. interdigitale</i> ATCC®9533	Growth after 3-7 days
<i>E. coli</i> ATCC®25922	Growth after 24 hours

Microbial testing was performed aerobically at 20-25°C

Sterility testing was performed at 35-37°C for 24 hours


Growth approved by: _____


Microbial Load

Lot Size:	54L
Number of Units Sampled:	14
Final pH at 25°C:	5.6±0.2

Appearance

Uninoculated broth is light to dark amber and opalescent

Lot approved by: _____


(QA Manager)

This product is not intended for the diagnosis or treatment of human disease