



# SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

## Certificate of Analysis Symphony Agar

Product Number: 16-Symph  
Lot Number: 2604280

Expiration: 5/28/2026

Fill Volume: 55mL

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

### Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Reaction</u>
<i>C. albicans</i> ATCC 10231	Growth	Round, pink colonies
<i>A. brasiliensis</i> ATCC 16404	Growth	Irregular, green colonies
<i>E. coli</i> ATCC 25922	Inhibited	N/A

Microbial testing was performed aerobically at 35-37°C for 18-24 hours  
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by: \_\_\_\_\_

  
(QA Manager)

### Microbial Load

Lot Size: 0.8L  
Number of Units Sampled: 30  
Final pH at 25°C: 5.6±0.2

### Appearance

Limpid, violet agar

Lot approved by: \_\_\_\_\_

  
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.