



SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

TCBS Agar

Product Number: 11-TCBS0
Lot Number: 2605180

Expiration: 8/18/2026

Fill Volume: 20 ± 2 mL
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>
<i>E. coli</i> ATCC® 25922	Inhibited
<i>S. typhimurium</i> ATCC® 14208	Inhibited
<i>S. flexneri</i> ATCC® 12022	Inhibited

Microbial testing was performed aerobically at 35°C for 18-24 hours

Growth approved by: _____
(QA Manager)

Microbial Load

Lot Size: 1L
Number of Units Sampled: 30
Final pH at 25°C: 8.6±0.2

Appearance

Uninoculated media is medium blue and clear

Lot approved by: _____
(QA Manager)

This product is not intended for diagnosis or treatment of human disease.