SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Modified Letheen Agar with Tween 80

Product Number: 11-MLAT0 Expiration: 9/15/2026 Fill Volume: 20 mL
Lot Number: 2606150 Fill Depth 3-5mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
E. coli ATCC® 25922	Growth
S. aureus ATCC® 6538	Growth
C. albicans ATCC® 10231	Growth
A. brasiliensis ATCC® 16404	Growth
B. subtilis ATCC® 6633	Growth
P. aeruginosa ATCC® 9027	Growth

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____
(QA Manager)

Microbial Load

Lot Size:	5L
Number of Units Sampled:	20
Final pH at 25°C:	7.2±0.2

Appearance

tannish yellow

Lot approved by:_____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.