SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Washed Sheep Blood Agar with Mitomycin C and Calcium Chloride

Product Number: 11-WSBAM
Lot Number: 2607060
Expiration: 10-6-26

Fill Volume: 20 ± 2 mL
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth	Reaction
E. coli ATCC®25922	Excellent	Hemolysis at 6 hours
E. coli ATCC®35150	Excellent	Hemolysis at 24 hours

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____
(QA Manager)

Microbial Load

Lot Size:	2L
Number of Units Sampled:	30
Final pH at 25°C:	7.2±0.2

Appearance

Uninoculated Washed Sheep Blood Agar is bright to dark red and opaque.

Lot approved by:_____
(QA Manager)

This product is not intended for *in vitro* diagnostic use