SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Mueller Hinton Agar

Product Number: 16-MHA00 Expiration: 11/6/2026 Fill Volume: 60 mL
Lot Number: 2607060

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth

**Organism recommended by CLSI M02-A*

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____
(QA Manager)

Microbial Load

Lot Size:	44L
Number of Units Sampled:	30
Final pH at 25°C:	7.3±0.2

Appearance

Lot approved by:_____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.