SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

R2A Agar Tubes

Product Number: 24-R2A20 Fill Volume: 20 ± 1 mL

Lot Number: 2606240 Expiration: 12/24/2026

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
E. faecalis ATCC®29212	Growth
E. coli ATCC®25922	Growth
K. aerogenes ATCC®13048	Growth
S. aureus ATCC®25923	Growth

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____

(QA Manager)

Microbial Load

Lot Size:	4.5
Number of Units Sampled:	12
Final pH at 25°C:	7.2±0.2

Appearance

Prepared R2A Agar is clear and light amber

Lot approved by:_____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.