SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Tryptic Soy Broth (Single)

Product Number: 24-TSBS25
Lot Number: 2606240

Fill Volume: 25 ± 1 mL
Expiration: 12/24/2026

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
E. coli ATCC® 25922	Good
E. coli ATCC®9027	Good
S. aureus ATCC® 25923	Good
S. aureus ATCC®6538	Good
B. subtilis ATCC®6633	Good
P. aeruginosa ATCC®9027	Good

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____ _____
(QA Manager)

Microbial Load

Lot Size:	1.1
Number of Units Sampled:	20
Final pH at 25°C:	7.3±0.2

Appearance

Uninoculated media is a light amber color and clear with possible precipitate

Lot approved by:_____
(QA Manager)

This product is not intended for the diagnoses or treatment of human disease.