SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Modified Letheen Broth

Product Number: XX-LeBM45 Fill Volume: 45 mL±2 mL

Lot Number: 2606110 Expiration: 12/11/2026

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
E. coli ATCC® 25922	Excellent
S. aureus ATCC® 25923	Excellent
C. albicans ATCC® 10231	Excellent

Microbial testing was performed anaerobically at 35-37°C for 18-24 hours

Sterility check was performed at 35-37°C for 24 hours

Growth approved by:_____ _____

(QA Manager)

Microbial Load

Lot Size:	3L
Number of Units Sampled:	20
Final pH at 25°C:	7.2 ± 0.2

Appearance

Prepared broth is straw colored, hazy, and sudsy when shaken.

Lot approved by:_____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.